

1 **EU - TYPE EXAMINATION CERTIFICATE**

2 **Equipment or Protective System Intended for use in Potentially Explosive Atmospheres**
Directive 2014/34/EU

3 EU - Type Examination Certificate **Baseefa08ATEX0086X – Issue 8**
Number:

3.1 In accordance with Article 41 of Directive 2014/34/EU, EC-Type Examination Certificates referring to 94/9/EC that were in existence prior to the date of application of 2014/34/EU (20 April 2016) may be referenced as if they were issued in accordance with Directive 2014/34/EU. Supplementary Certificates to such EC-Type Examination Certificates, and new issues of such certificates, may continue to bear the original certificate number issued prior to 20 April 2016.

4 Product: **HS-420 Series Accelerometer**

5 Manufacturer: **Hansford Sensors Limited**

6 Address: **Sands Industrial Estate, Bucks, HP12 4HJ**

7 This re-issued certificate extends EC Type Examination Certificate No. Baseefa08ATEX0086X to apply to product designed and constructed in accordance with the specification set out in the Schedule of the said certificate but having any variations specified in the Schedule attached to this certificate and the documents therein referred to.

8 SGS Baseefa, Notified Body number 1180, in accordance with Article 17 of Directive 2014/34/EU of the European Parliament and of the Council, dated 26 February 2014, certifies that this product has been found to comply with the Essential Health and Safety Requirements relating to the design and construction of products intended for use in potentially explosive atmospheres given in Annex II to the Directive.

The examination and test results are recorded in confidential Report No. See Certificate History

9 Compliance with the Essential Health and Safety Requirements has been assured by compliance with:

EN 60079-0: 2012 + A11: 2013 EN 60079-11: 2012

except in respect of those requirements listed at item 18 of the Schedule.

10 If the sign “X” is placed after the certificate number, it indicates that the product is subject to the Specific Conditions of Use specified in the schedule to this certificate.

11 This EU - TYPE EXAMINATION CERTIFICATE relates only to the design and construction of the specified product. Further requirements of the Directive apply to the manufacturing process and supply of this product. These are not covered by this certificate.

12 The marking of the product shall include the following :

⊕ II 1G Ex ia IIC T6 Ga (-40°C ≤ T_a ≤ +60°C)

⊕ II 1D Ex ia IIIC T80°C IP65 Da (-40°C ≤ T_a ≤ +60°C)

⊕ I M1 Ex ia I Ma (-40°C ≤ T_a ≤ +60°C)

SGS Baseefa Customer Reference No. **5943**

Project File No. **17/0352**

This document is issued by the Company subject to its General Conditions for Certification Services accessible at <http://www.sgs.com/en/Terms-and-Conditions.aspx> and the Supplementary Terms and Conditions accessible at <http://www.sgs.com/SGSBaseefa/Terms-and-Conditions.aspx>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained herein reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. It does not necessarily indicate that the equipment may be used in particular industries or circumstances. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, schedule included, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law.

SGS Baseefa Limited

Rockhead Business Park, Staden Lane,
Buxton, Derbyshire SK17 9RZ

Telephone +44 (0) 1298 766600 Fax +44 (0) 1298 766601

e-mail baseefa@sgs.com web site www.sgs.co.uk/baseefa

Registered in England No. 4305578.

Registered address: Rossmore Business Park, Ellesmere Port, Cheshire, CH65 3EN

R S Sinclair
R S SINCLAIR PP DOREADLEY

TECHNICAL MANAGER

On behalf of SGS Baseefa Limited

13

Schedule

14

Certificate Number Baseefa08ATEX0086X – Issue 8

15 Description of Product

The HS-420 Series Accelerometers are designed to measure velocity or acceleration by converting the signal generated by the compression of a piezo electric crystal by a given seismic mass and output a 4 to 20mA signal proportional to velocity or acceleration to the monitoring equipment.

The accelerometer comprises a piezo electric crystal connected to a signal conditioning board all contained within a stainless steel enclosure of various shapes measuring approximately 33cm³. The enclosure is a fully welded construction.

Electrical connections are made to the apparatus either via an IP65 rated connector or via an integral cable which is encapsulated in the end of the apparatus.

The Group II version of the apparatus (excluding cable) has the following terminal parameters:

$$\begin{aligned} U_i &= 28V & C_i &= 0 \\ I_i &= 115mA & L_i &= 0 \\ P_i &= 0.65W \end{aligned}$$

The HS-420IT or HS-422IT Accelerometer, with an integral temperature sensor and 25m of cable has the following terminal parameters for all four electrical connections considered as a single intrinsically safe circuit:

$$\begin{aligned} U_i &= 44V & C_i &= 3nF \\ I_i &= 117mA & L_i/R_i &= 13nH/\Omega \\ P_i &= 0.722W \end{aligned}$$

The Group I version of the apparatus (excluding cable) has the following terminal parameters:

$$\begin{aligned} U_i &= 16.5V & C_i &= 0 \\ P_i &= 1.74W & L_i &= 0 \end{aligned}$$

The apparatus must be powered from a power limited source such as an appropriately certified fuse assembly containing a $\leq 62mA$ fuse, 1.74W (16.5V x 62mA x 1.7).

The Group I version of the apparatus (excluding cable) has the following alternative terminal parameters:

$$\begin{aligned} U_i &= 28V & C_i &= 0 \\ I_i &= 115mA & L_i &= 0 \\ P_i &= 0.65W \end{aligned}$$

The capacitance and inductance to resistance ratio of the different versions have the following parameters:

	Integral Cable or 2-Pin Mill Spec Connector with cable				4-Pin M12 Connector
	Polyurethane Cable	Silicone Cable	Armoured Cable	PUR Cable	Polyurethane Cable
C_i	= 160pF/m	= 370pF/m	= 290pF/m	= 884pF/m	= 120pF/m
L_i	= 0.72μH/m	= 0.5μH/m	= 0.5μH/m	= 0.6μH/m	= 0.7μH/m
L_i/R_i	= 8.32μH/Ω	= 15.4μH/Ω	= 15.4μH/Ω	= 6.1μH/Ω	= 11.7μH/Ω

16 Report Number

See Certificate History

17 Specific Conditions of Use

1. The free end of the cable on the integral cable version of the apparatus must be terminated in an appropriately certified dust proof enclosure

18 Essential Health and Safety Requirements

In addition to the Essential Health and Safety Requirements (EHSRs) covered by the standards listed at item 9, the following are considered relevant to this product:

Clause	Subject	Compliance
1.2.7	LVD type requirements	Standards require manufacturer's declaration, supplied.
1.2.8	Overloading of equipment (protection relays, etc.)	Covered by installation rules and manufacturer's instructions
1.4.1	External effects	The Purchaser should make the manufacturer aware of such issues. Covered in Instructions
1.4.2	Aggressive substances, etc.	The Purchaser should make the manufacturer aware of such issues. Covered in Instructions

19 Drawings and Documents

New drawings submitted for this issue of certificate:

None

Current drawings which remain unaffected by this issue:

Number	Sheet	Issue	Date	Description
M06-008-F	1 to 5	F	12/06/15	General Arrangement and Product Information for Group I, Group II and Group III HS-420 & HS-422 Series Accelerometers
M06-009-B	1 of 1	B	13/01/11	Zener Diode Arrangement HS-420 & HS-422 Series Accelerometer
M06-014-A	1 of 2	A	19.03.08	Din Rail Mounted Enclosure with a 62mA Safe-T-Fuse 259 Series F.U.W. HS-420M/HS422M Series Group I Accelerometers
M06-014-A	2 of 2	A	19.03.08	Inline Fuse with a 62mA Pico II 251 Series Fuse F.U.W. HS-420M/HS422M Series Group I Accelerometers
M06-020-C	1 to 4	C	30/08/12	General Arrangement and Product Information for Group I Fused HS-420 & HS-422 Series Accelerometer
M06-022-C	1 to 3 of 4	C	30/08/12	General Arrangement and Product Information for Group II HS-420IT & HS-422IT Series Accelerometer
M06-022-C	4 of 4	C	30/08/12	General Arrangement and Product Information for Group II and Group III HS-420IT & HS-422IT Series Accelerometer
HS420M	1 of 1	C	14.01.11	HS-420M/HS-422M 4-20mA Circuit
HS420IT	1 of 1	D	14.01.11	HS-420IT/HS-422IT 4-20mA Circuit
HS420-IS	1 of 1	D	21/07/15	HS-420/HS-422 4-20mA Circuit
P01-004	1 of 1	D	05.01.11	4-20mA PCB Track Layout
P02-004	1 of 1	D	05.01.11	4-20mA PCB Component Layout
P01-012	1 of 1	A	03.03.08	HS420I PCB for M12 Connector
P02-012	1 of 1	A	27.02.08	HS420I PCB Connector
P01-026	1 of 1	A	19.05.08	HS420M Connection PCB for M12 Connector
P02-026	1 of 1	A	19.05.08	HS420M Connection PCB
P01-027	1 of 1	B	09.09.08	HS420IT Connection PCB for M12 Connector
P02-027	1 of 1	B	09.09.08	HS420IT Connection PCB

Number	Sheet	Issue	Date	Description
P01.080	1 of 1	A	19/05/15	HS420I PCB for 2-Pin Connector
P02-080	1 of 1	A	19/05/15	HS420I PCB 2-Pin MS Connector Component Layout

20 Certificate History

Certificate No.	Date	Comments
Baseefa08ATEX0086X	10 April 2008	The release of the prime certificate. The associated test and assessment is documented in Test Report GB/BAS/ExTR08.0059/00.
Baseefa08ATEX0086/1	1 July 2008	To permit the introduction of a new mining version complete with integral fuse identified as: HS-420MFxxxxyyzz or HS-422MFxxxxyyzz. To permit minor non-electrical changes that do not affect the original assessment. Intrinsic safety is not affected. The associated test and assessment is documented in Test Report GB/BAS/ExTR08.0112/00.
Baseefa08ATEX0086/2	18 September 2008	To permit the introduction the HS-420IT or HS-422IT Accelerometer, with an integral temperature sensor, 25m of cable and associated electrical connections. Intrinsic safety is not affected. The associated test and assessment is documented in Test Report GB/BAS/ExTR08.0181/00.
Baseefa08ATEX0086/3	4 February 2009	To permit minor mechanical changes that do not affect the original assessment. Intrinsic safety is not affected. The associated test and assessment is documented in Test Report GB/BAS/ExTR09.0014/00.
Baseefa08ATEX0086/4	24 January 2011	To permit minor electrical changes that do not affect the original assessment. To permit minor marking changes that do not affect the original assessment. Clarification of components not fitted to PCB added to report. Intrinsic safety is not affected. The associated test and assessment is documented in Test Report GB/BAS/ExTR11.0013/00.
Baseefa08ATEX0086/5	2 February 2012	To permit the HS-420M series of the equipment (not the HS-420MF series) to be marked with alternative terminal parameters. Intrinsic safety is not affected. The associated test and assessment is documented in Test Report GB/BAS/ExTR12.0005/00.
Baseefa08ATEX0086/6	10 October 2012	To confirm that the equipment covered by this certificate also meets the requirements of EN60079-0: 2012 (IEC60079-0: 2011, Edition 6) and EN60079-11: 2012 (IEC60079-11: 2011, Edition 6) as supported by GB/BAS/ExTR12.0254/00 held on technical file IECEx BAS 07.0035X.
Baseefa08ATEX0086X Issue 7	17 June 2015	This issue of the certificate incorporates previously issued primary & supplementary certificates into one certificate and permits the following changes: To permit the use of an alternative PUR cable. To permit the use of an alternative 2-Pin connector. Intrinsic safety is not affected. The associated test and assessment is documented in Test Report GB/BAS/ExTR15.0175/00. Re-Issued 22 July 2015 to include references to 2-pin connector drawings and replace original. The associated test and assessment is documented in Test Report GB/BAS/ExTR15.0175/01.

Certificate Number
Baseefa08ATEX0086X
Issue 8



Issued 16 May 2017
Page 5 of 5

Certificate No.	Date	Comments
Baseefa08ATEX0086X Issue 8	16 May 2017	Addition of C _i & L _i within the entity parameter information and confirmation that the equipment complies with the requirements of the Directive 2014/34/EU.
For drawings applicable to each issue, see original of that issue.		