



IECEx Certificate of Conformity

INTERNATIONAL ELECTROTECHNICAL COMMISSION IEC Certification Scheme for Explosive Atmospheres

for rules and details of the IECEx Scheme visit www.iecex.com

Certificate No.: IECEx BAS 18.0082X

Issue No: 0

Certificate history:

[Issue No. 0 \(2018-11-09\)](#)

Status: **Current**

Page 1 of 5

Date of Issue: **2018-11-09**

Applicant: **Hansford Sensors Limited**

Artisan
Hillbottom Road
Sands Industrial Estate
Bucks
HP12 4HJ
United Kingdom

Equipment: **HS-150I & HS170I Series Accelerometers**

Optional accessory:

Type of Protection: **Intrinsic Safety**

Marking:
See Certificate Schedule for Marking Details

Approved for issue on behalf of the IECEx
Certification Body:

R. Sinclair


D BREARLEY
Certification
Manager

Position:

Technical Manager

Signature:
(for printed version)

Date:


9/11/18

1. This certificate and schedule may only be reproduced in full.
2. This certificate is not transferable and remains the property of the issuing body.
3. The Status and authenticity of this certificate may be verified by visiting the [Official IECEx Website](#).

Certificate issued by:

SGS Baseefa Limited
Rockhead Business Park
Staden Lane
Buxton, Derbyshire, SK17 9RZ
United Kingdom





IECEx Certificate of Conformity

Certificate No: IECEx BAS 18.0082X Issue No: 0

Date of Issue: 2018-11-09 Page 2 of 5

Manufacturer: **Hansford Sensors Limited.**
Artisan
Hillbottom Road
Sands Industrial Estate
Bucks
HP12 4HJ
United Kingdom

Additional Manufacturing location(s):

This certificate is issued as verification that a sample(s), representative of production, was assessed and tested and found to comply with the IEC Standard list below and that the manufacturer's quality system, relating to the Ex products covered by this certificate, was assessed and found to comply with the IECEx Quality system requirements. This certificate is granted subject to the conditions as set out in IECEx Scheme Rules, IECEx 02 and Operational Documents as amended.

STANDARDS:

The apparatus and any acceptable variations to it specified in the schedule of this certificate and the identified documents, was found to comply with the following standards:

IEC 60079-0 : 2017 Edition:7.0	Explosive atmospheres - Part 0: Equipment - General requirements
IEC 60079-11 : 2011 Edition:6.0	Explosive atmospheres - Part 11: Equipment protection by intrinsic safety "I"

*This Certificate **does not** indicate compliance with electrical safety and performance requirements other than those expressly included in the Standards listed above.*

TEST & ASSESSMENT REPORTS:

A sample(s) of the equipment listed has successfully met the examination and test requirements as recorded in

Test Report:

[GB/BAS/ExTR18.0244/00](#)

Quality Assessment Report:

[GB/BAS/QAR07.0040/07](#)



IECEx Certificate of Conformity

Certificate No: IECEx BAS 18.0082X

Issue No: 0

Date of Issue: 2018-11-09

Page 3 of 5

Schedule

EQUIPMENT:

Equipment and systems covered by this certificate are as follows:

The HS-150i and HS-170i Series Accelerometers are designed to measure acceleration or vibration by converting the signal generated by the compression of a Piezo electric crystal by a given seismic mass and outputting a broadband ac signal to the monitoring equipment.

The accelerometer comprises of a piezo electric crystal connected to a signal conditioning board, all contained within a fully welded steel enclosure.

HS-150xT versions include a temperature transmitter.

HS-173 is a tri-axial accelerometer comprising three individual circuits with common 0V line, sharing a single set of parameters.

Electrical connections are made via a connector or integral cable.

The equipment carries the following markings:

Uni-axial Accelerometer, Integral Cable Version

Ex ia IIC T6...T4 Ga
Ex ia IIIC T110°C... T145°C Da
Ex ia I Ma

Uniaxial Accelerometer, Connector

Ex ia IIC T6...T4 Ga
Ex ia IIIC T135°C Da
Ex ia IIIB T110°C... T145°C Da
Ex ia I Ma

Triaxial Accelerometer – Integral cable

Ex ia IIC T6...T4 Ga
Ex ia IIIC T102°C... T131°C Da
Ex ia I Ma

Triaxial Accelerometer – Connector

Ex ia IIC T6...T4 Ga
Ex ia IIIC T135°C Da
Ex ia IIIB T102°C... T131°C Da
Ex ia I Ma

The equipment has the following temperature parameters:

Uni-axial accelerometers with integral cable

Ex ia IIC T6 Ga -55°C ≤ Ta ≤ +57°C
Ex ia IIC T4 Ga -55°C ≤ Ta ≤ +103°C

Ex ia IIIC T₂₀₀110°C Da -55°C ≤ Ta ≤ +57°C
Ex ia IIIC T₂₀₀135°C Da -55°C ≤ Ta ≤ +70°C
Ex ia IIIC T₂₀₀145°C Da -55°C ≤ Ta ≤ +92°C

Ex ia I Ma -55°C ≤ Ta ≤ +103°C

Uni-axial accelerometers with connectors

Ex ia IIC T6 Ga -55°C ≤ Ta ≤ +57°C
Ex ia IIC T4 Ga -55°C ≤ Ta ≤ +103°C



IECEx Certificate of Conformity

Certificate No: IECEx BAS 18.0082X

Issue No: 0

Date of Issue: 2018-11-09

Page 4 of 5

Ex ia IIIC T₂₀₀ 135°C Da -55°C ≤ Ta ≤ +70°C

Ex ia IIIB T₂₀₀ 110°C Da -55°C ≤ Ta ≤ +57°C

Ex ia IIIB T₂₀₀ 145°C Da -55°C ≤ Ta ≤ +92°C

Ex ia I Ma -55°C ≤ Ta ≤ +103°C

Tri-axial Accelerometers with integral cable.

Ex ia IIC T6 Ga -55°C ≤ Ta ≤ +69°C

Ex ia IIC T4 Ga -55°C ≤ Ta ≤ +104°C

Ex ia IIIC T₂₀₀ 102°C Da -55°C ≤ Ta ≤ +69°C

Ex ia IIIC T₂₀₀ 131°C Da -55°C ≤ Ta ≤ +98°C

Ex ia I Ma -55°C ≤ Ta ≤ +104°C

Triaxial accelerometers with connectors.

Ex ia IIC T6 Ga -55°C ≤ Ta ≤ +69°C

Ex ia IIC T4 Ga -55°C ≤ Ta ≤ +104°C

Ex ia IIIC T₂₀₀ 135°C Da -55°C ≤ Ta ≤ +70°C

Ex ia IIIB T₂₀₀ 102°C Da -55°C ≤ Ta ≤ +69°C

Ex ia IIIB T₂₀₀ 131°C Da -55°C ≤ Ta ≤ +98°C

Ex ia I Ma -55°C ≤ Ta ≤ +104°C

See Certificate Annex for terminal parameters

SPECIFIC CONDITIONS OF USE: YES as shown below:

1. Where equipment is supplied with an attached cable, this must be terminated in an enclosure providing at least degree of protection IP20.
2. The equipment is marked with reduced certification markings. Refer to the Certificate Schedule for the full certification markings & applicable temperature classification and associated ambient temperature range.



IECEX Certificate of Conformity

Certificate No: IECEx BAS 18.0082X

Issue No: 0

Date of Issue: 2018-11-09

Page 5 of 5

Annex:

[IECEX BAS 18.0082X Annex Iss 0.pdf](#)

The HS-150i and HS-170i Series Accelerometers

The equipment has the following terminal parameters:

Uni-axial accelerometer.

Connector Only		
Ui	=	28V
Ii	=	93mA
Pi	=	0.65W
Ci	=	1.2nF
Li	=	0

10m of cable		
Ui	=	28V
Ii	=	93mA
Pi	=	0.65W
Ci	=	5.0nF
Li	=	7.2μH

92m of cable		
Ui	=	28V
Ii	=	93mA
Pi	=	0.65W
Ci	=	35.9nF
Li	=	66μH

Tri-axial accelerometer

Connector Only		
Ui	=	28V
Ii	=	93mA
Pi	=	0.65W
Ci	=	3.6nF
Li	=	0

10m of cable		
Ui	=	28V
Ii	=	93mA
Pi	=	0.65W
Ci	=	7.4nF
Li	=	7.2μH

92m of cable		
Ui	=	28V
Ii	=	93mA
Pi	=	0.65W
Ci	=	38.3nF
Li	=	66μH