



IECEx Certificate of Conformity

INTERNATIONAL ELECTROTECHNICAL COMMISSION IEC Certification System for Explosive Atmospheres

for rules and details of the IECEx Scheme visit www.iecex.com

Certificate No.: **IECEx BAS 08.0034X**

Page 1 of 4

Status: **Current**

Issue No: 14

Date of Issue: 2021-04-13

Applicant: **Hansford Sensors Limited**
Artisan
Hillbottom Road
Sands Industrial Estate
Bucks
HP12 4HJ
United Kingdom

Certificate history:

Issue 13 (2020-09-10)
Issue 12 (2018-08-17)
Issue 11 (2017-11-01)
Issue 10 (2017-05-16)
Issue 9 (2015-06-17)
Issue 8 (2013-01-15)
Issue 7 (2012-10-12)
Issue 6 (2012-02-02)
Issue 5 (2011-01-24)
Issue 4 (2009-11-30)

Equipment: **HS-420 Series Accelerometer**

Optional accessory: Fuse Box

Type of Protection: **Intrinsic Safety, gas and dust**

Marking: **Ex ia I Ma (-40°C ≤ Ta ≤ +60°C)**

Ex ia IIC T4/T6 Ga (-40°C ≤ Ta ≤ +110°C/60°C)

Ex ia IIIC T130°C/T80°C IP65 Da (-40°C ≤ Ta ≤ +110°C/60°C)

Approved for issue on behalf of the IECEx
Certification Body:

R S Sinclair

Position:

Technical Manager

Signature:
(for printed version)

Date:

14.4.2021

1. This certificate and schedule may only be reproduced in full.
2. This certificate is not transferable and remains the property of the issuing body.
3. The Status and authenticity of this certificate may be verified by visiting www.iecex.com or use of this QR Code.



Certificate issued by:

SGS Baseefa Limited
Rockhead Business Park
Staden Lane
Buxton, Derbyshire, SK17 9RZ
United Kingdom





IECEx Certificate of Conformity

Certificate No.: **IECEx BAS 08.0034X**

Page 2 of 4

Date of issue: 2021-04-13

Issue No: 14

Manufacturer: **Hansford Sensors Limited**
Artisan
Hillbottom Road
Sands Industrial Estate
Bucks
HP12 4HJ
United Kingdom

Additional
manufacturing
locations:

This certificate is issued as verification that a sample(s), representative of production, was assessed and tested and found to comply with the IEC Standard list below and that the manufacturer's quality system, relating to the Ex products covered by this certificate, was assessed and found to comply with the IECEx Quality system requirements. This certificate is granted subject to the conditions as set out in IECEx Scheme Rules, IECEx 02 and Operational Documents as amended

STANDARDS :

The equipment and any acceptable variations to it specified in the schedule of this certificate and the identified documents, was found to comply with the following standards

IEC 60079-0:2011 Explosive atmospheres - Part 0: General requirements
Edition:6.0

IEC 60079-11:2011 Explosive atmospheres - Part 11: Equipment protection by intrinsic safety "i"
Edition:6.0

This Certificate **does not** indicate compliance with safety and performance requirements other than those expressly included in the Standards listed above.

TEST & ASSESSMENT REPORTS:

A sample(s) of the equipment listed has successfully met the examination and test requirements as recorded in:

Test Reports:

[GB/BAS/ExTR08.0059/00](#)
[GB/BAS/ExTR09.0014/00](#)
[GB/BAS/ExTR12.0254/00](#)
[GB/BAS/ExTR18.0140/00](#)

[GB/BAS/ExTR08.0112/00](#)
[GB/BAS/ExTR11.0013/00](#)
[GB/BAS/ExTR15.0175/00](#)
[GB/BAS/ExTR20.0138/00](#)

[GB/BAS/ExTR08.0181/00](#)
[GB/BAS/ExTR12.0005/00](#)
[GB/BAS/ExTR17.0322/00](#)
[GB/BAS/ExTR21.0054/00](#)

Quality Assessment Report:

[GB/BAS/QAR07.0040/09](#)



IECEx Certificate of Conformity

Certificate No.: **IECEx BAS 08.0034X**

Page 3 of 4

Date of issue: 2021-04-13

Issue No: 14

EQUIPMENT:

Equipment and systems covered by this Certificate are as follows:

The HS-420 Series Accelerometers are designed to measure velocity or acceleration by converting the signal generated by the compression of a piezo electric crystal by a given seismic mass and output a 4 to 20mA signal proportional to velocity or acceleration to the monitoring equipment.

The accelerometer comprises a piezo electric crystal connected to a signal conditioning board all contained within a stainless-steel enclosure of various shapes measuring approximately 33cm³. The enclosure is a fully welded construction.

Electrical connections are made to the apparatus either via an IP65 rated connector or via an integral cable which is encapsulated in the end of the apparatus.

The HS-420 & HS-422 Series Accelerometers are designed to measure velocity and acceleration respectively and output a 4 to 20mA output. Variants of these series of accelerometer are also available with an integral temperature sensor with Model No's HS-420IT & HS-422IT. The HS-421 & HS423 Series Accelerometers are similar to the HS-420 & HS-422 Series measuring velocity and acceleration respectively, but are fitted with an additional AC acceleration output.

For terminal parameters see Annex

SPECIFIC CONDITIONS OF USE: YES as shown below:

1. The free end of the cable on the integral cable version of the apparatus must be terminated in an appropriately certified dust proof enclosure.



IECEx Certificate of Conformity

Certificate No.: **IECEx BAS 08.0034X**

Page 4 of 4

Date of issue: 2021-04-13

Issue No: 14

DETAILS OF CERTIFICATE CHANGES (for issues 1 and above)

Variation 14.1

To permit the introduction the HS-421 & HS-423 Series of Accelerometers. These series of accelerometers are similar in construction to the HS-420 & HS-422I Accelerometer, but are fitted with an additional AC output. The Certificate Schedule was revised to include the details of the new models and the Certificate Annex (Now Issue 5) updated.

ExTR: GB/BAS/ExTR21.0054/00	File Reference: 20/0588
------------------------------------	--------------------------------

Annex:

[IECEx BAS 08.0034X Annex Issue 5.pdf](#)

HS-420 Series Accelerometer

Input Parameters

The Group II & III versions of the HS-420, HS-421, HS-422 & HS-423 Series Accelerometers (excluding cable) have the following terminal parameters:

$$\begin{aligned} U_i &= 28V & C_i &= 0 \\ I_i &= 115mA & L_i &= 0 \\ P_i &= 0.65W \end{aligned}$$

The HS-420IT or HS-422IT Accelerometer, with an integral temperature sensor and 25m of cable has the following terminal parameters for all four electrical connections considered as a single intrinsically safe circuit:

$$\begin{aligned} U_i &= 44V & C_i &= 3nF \\ I_i &= 117mA & L_i/R_i &= 13nH/\Omega \\ P_i &= 0.722W \end{aligned}$$

The Group I versions of the HS-420 & HS-422 Series Accelerometers (excluding cable) have the following terminal parameters:

$$\begin{aligned} U_i &= 16.5V & C_i &= 0 \\ P_i &= 1.74W & L_i &= 0 \end{aligned}$$

The apparatus must be powered from a power limited source such as an appropriately certified fuse assembly containing a $\leq 62mA$ fuse, 1.74W (16.5V x 62mA x 1.7).

The Group I version of the apparatus (excluding cable) has the following alternative terminal parameters:

$$\begin{aligned} U_i &= 28V & C_i &= 0 \\ I_i &= 115mA & L_i &= 0 \\ P_i &= 0.65W \end{aligned}$$

The capacitance and inductance to resistance ratio of the different versions have the following parameters:

	Integral Cable or 2-Pin Mil Spec Connector with cable				4-Pin M12 Connector
	Polyurethane Cable	Silicone Cable	Armoured Cable	PUR Cable	Polyurethane Cable
C_i	= 160pF/m	= 370pF/m	= 290pF/m	= 884pF/m	= 120pF/m
L_i	= 0.72 μ H/m	= 0.5 μ H/m	= 0.5 μ H/m	= 0.6 μ H/m	= 0.7 μ H/m
L_i/R_i	= 8.32 μ H/ Ω	= 15.4 μ H/ Ω	= 15.4 μ H/ Ω	= 6.1 μ H/ Ω	= 11.7 μ H/ Ω

The above equipment is marked with the following certification markings and associated ambient temperature ranges:

Group I Accelerometers Models HS-420M, HS-420MF, HS-422M & HS-422MF:	Ex ia I Ma (-40°C \leq T _a \leq +60°C)
Group II & III Accelerometers Models HS-420I, HS-420IT, HS-421I, HS-422I, HS-422IT & HS-423I:	Ex ia IIC T4 Ga (-40°C \leq T _a \leq +110°C) Ex ia IIIC T130°C IP65 Da (-40°C \leq T _a \leq +110°C) Ex ia IIC T6 Ga (-40°C \leq T _a \leq +60°C) Ex ia IIIC T80°C IP65 Da (-40°C \leq T _a \leq +60°C)