

Connection Details of Cable Assembly





Connector sealed to IP66 only when

connected to the accelerometer.









Materials of Cable: C01-025 Polyurethane Screened Coiled Cable			
Pos	Description	Overall Diameter	Remarks
1.	Tin plated copper conductor	1.02	24 AWG
2.	Screen wire of tin plated copper		24 AWG
3.	Poly filler insulation		
4.	Paper wrap insulation seperator		
5.	Braided screen of tin plated copper		
6.	Paper wrap insulation seperator		
7.	Polyurethane jacket	5.4mm (±0.1mm)	Non flame retardant, rated from -40°C to +80°C



2 Pin MS Overmoulded

Booted Straight Socket with Stainless Steel

Locking Ring

Tel: +44(0) 845 680 1957 Fax: +44(0) 845 680 1958 Email: sales@hansfordsensors.com Web: www.hansfordsensors.com