
# HS-150tb Premium Accelerometer AC acceleration output via PUR Cable with Protective Conduit



Technical Performance	;	Mechanical	
Mounted Base Resonance	see 'How To Order' table (nominal)	Case Material	Stainless Steel
Sensitivity	see: 'How To Order' table ±10%	Sensing Element/Construction	PZT/Shear
	Nominal 80Hz at 22°C	Mounting Torque	8Nm
Frequency Response	2Hz (120cpm) to 10kHz (600kcpm) ± 5%	Weight	106gms (nominal) body only
	1.5Hz (90cpm) to 12kHz (720kcpm) ± 10%	Maximum Cable Length	See certificate
	0.8Hz (48cpm) to 15kHz (900kcpm) ± 3dB	Standard Cable Length	5 metres
Isolation	Base isolated	Screened Cable	PUR - length to be specified with order
Range	see: 'How To Order' table	Mounting Threads	see: 'How To Order' table
Transverse Sensitivity	Less than 5%	Submersible Depth	100 metres max (10 bar)

Electrical	
Excitation Voltage:	18-30Volts DC
Electrical Noise	0.1mg max
Current Range	0.5mA to 8mA
Bias Voltage	10 - 12 Volts DC
Settling Time	2 seconds
Output Impedance	200 Ohms max.
Case Isolation	>10 <sup>8</sup> Ohms at 500 Volts

### Typical Frequency Response (at 100mV/g)



## Applications

Environmental

Sealing Maximum Shock

EMC

**Operating Temperature Range** 

Fans, Motors, Pumps, Compressors, Centrifuges, Conveyors, Air Handlers, Gearboxes, Rolls, Dryers, Presses, Cooling, VAC, Spindles, Machine Tooling, Process Equipment

Vibration sensor should be firmly fixed to a flat surface (spot face surface may be needed to be produced and cable anchored to sensor body.)



see: attached certification details

IP68

5000g

EN61326-1:2013

#### Certifications

Industries







www.hansfordsensors.com sales@hansfordsensors.com




We reserve the right to alter the specification of this product without prior notice TS1414.3

# **HS-150tb Premium Accelerometer**

AC acceleration output via PUR Cable with Protective Conduit

Ex to Requirements			
Certificate details: Group II	IECExEXV23.0048X	Certified Temperature Range	(-55°C ≤ Ta ≤ +82°C)
	ExVeritas23ATEX1702X		
	II 2 D Ex tb IIIC T110℃ Db		
500V Isolation	Units Will Pass A 500V Isolation Test		
Standards Applied to Product	EN IEC 60079-0:2018		
	EN 60079-31:2014		
	IEC 60079-0 Edition 7 2017		
	IEC 60079-31:2013		

#### How To Order





www.hansfordsensors.com sales@hansfordsensors.com



We reserve the right to alter the specification of this product without prior notice TS1414.3